



ACCREDITATION CERTIFICATE

004-LB-CAL

Emirates International Accreditation Centre

has accredited

GENERAL CONST. LAB CALIBRATION LLC

Industrial Area # 3, Sharjah-United Arab Emirates

In accordance with the requirements of

ISO/IEC 17025:2005

General requirements for the competence of testing and calibration laboratories

to undertake the calibration in the attached accreditation scope

This Accreditation is invalid without the attached accreditation scope and shall remain in force within the validity period printed below, subject to continuing compliance with the requirements of the accreditation criteria.

Validity: 25/12/2019 to 25/05/2021

Initial Accreditation Date: 25/05/2009



CHIEF EXECUTIVE OFFICER
APPROVAL



Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
All types of Caliper Digital/ Dial/ Vernier -0.02 mm	GTS-WP-22 Based on BS EN ISO 13385-1 & BS 887 (only for limits of error reference) For determining error of indicated size Comparison with gauge blocks	0 – 600 mm	30 µm	Laboratory/ Customer Premises (Std. room/ Metrology)
		> 600mm to 1500mm	50 µm	
		1500mm < to 2000mm	60 µm	
Vernier Caliper (0.05 mm)	GTS-WP-22 Based on BS EN ISO 13385- 1 & BS 887 (only for limits of error reference) For determining error of indicated size Comparison with gauge blocks	0 – 600 mm (already available)	36 µm	Laboratory/ Customer Premises (Std. room/ Metrology)
		600 mm < to 1500 mm	50 µm	
		1500 mm < to 2000 mm	60 µm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
External Micrometer (Digital) LC: 0.001 mm)	GTS-WP-23 Based on BS EN ISO 3611 & BS 870 (only for limits of error reference) DMS 2014 For determining error of indicated size Comparison with gauge blocks	0-25mm (already available)	2.5 µm	Laboratory/ Customer Premises (Std. room/ Metrology)
		>25 mm up to 100 mm	4 µm	
		>100 mm up to 500 mm	10 µm	
		>500 mm up to 925 mm	20 µm	
External Micrometer (Analogue LC:0.01 mm)	GTS-WP-23 Based on BS EN ISO 3611 & BS 870 (only for limits of error reference) For determining error of indicated size Comparison with gauge blocks	0 up to 25mm (already available)	3 µm	Laboratory/ Customer Premises (Std. room/ Metrology)
		>25mm up to 100mm	4 µm	
		>100mm up to 500mm	10 µm	
		>500mm up to 925mm	20 µm	
Micrometer Setting Standard	GTS-WP-23 Based on BS EN ISO 3611 For determining length using 1D comparator (ULMS)	Up to 100 mm	2 µm	Laboratory
		>100 up to 600 mm	9 µm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Micrometer Setting Standard	GTS-WP-23 Based on BS EN ISO 3611 (using HMS) For determining length using HMS	Up to 100mm	3 µm	Laboratory
		>100mm up to 600mm	10 µm	
		>600 mm up to 900 mm	15 µm	
Depth Micrometer (Digital/ analogue)	GTS-WP-28 Based on BS EN ISO 6468 For determining error of indicated depth Comparison with gauge blocks	Up to 25mm	3 µm	Laboratory
		>25 mm up to 100mm	4 µm	
		>100 mm up to 300 mm	6 µm	
		Up to 100mm /0.01mm	7 µm	
		>100mm up to 300mm /0.01 mm	12 µm	
Tubular Micrometer /Inside Micrometer (Digital/ Analogue)	GTS-WP-24 Based on BS EN ISO 959 For determining error of indicated size Comparison with gauge blocks and ULMS	Up to 300mm	5 µm	Laboratory
		>300mm up to 500mm	8 µm	
		>500 mm up to 1000 mm	15 µm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Tubular Micrometer /Inside Micrometer (Digital/ Analogue)	GTS-WP-24 Based on BS EN ISO 959 For determining error of indicated size Comparison with gauge blocks and ULMS	0.001 mm up to 300mm/0.01mm	8 µm	Laboratory
		300mm <up to 2000mm/ 0.01mm	30 µm	
Inside Micrometer (Caliper Type)	GTS-WP-24 Based on BS EN ISO 959 For determining error of indicated size Comparison with gauge blocks, ring gauges and ULMS	0 mm up to 50mm	3 µm	Laboratory/ Customer Premises (Std.
Dial/Digital Indicators	GTS-WP-26 Based on BS EN ISO 463& BS 907 (only for limits of error reference) For determining error of indicated displacement Comparison with ULMS	0.01mm up to 100mm /0.01mm	8 µm	
		0.001 mm up to 50 mm /0.001 mm	3 µm	
	Comparison with dial gauge calibrator	0.01mm up to 25mm /0.01mm	7 µm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Bore Gauge(Ordinary/ Digital)	GTS-WP-27 Based on JIS B 7515 For determining error of indicated diameter Comparison with calibration tester and ULMS	Up to 400 mm/0.001mm	9 µm	Laboratory
		Up to 400 mm/0.01mm	10 µm	
LVDT (Ordinary/ Digital)	GTS-WP-168 Based on ASTM F2537 For determining error of indicated displacement Mechanical comparison to calibrated gauge blocks/ULMS	UP to 200mm	3+(0.05*L) µm; L: mm	Laboratory/ Customer Premises (Std. room)
Dial Test Indicator/ Lever Type Dial Gauges	GTS-WP-172 Based on BS EN ISO 463, BS 2795 & IS 11498 For determining error of indicated displacement Comparison with ULMS	Up to 1mm/0.001mm	2 µm	Laboratory
		Up to 1mm/0.01mm	6 µm	
Dial Test Indicator/ Lever Type Dial Gauges	Comparison with dial gauge calibrator	Up to 1mm/0.01mm	7 µm	Laboratory/ Customer Premises

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Dial/Digital Thickness Gauges	GTS-WP-36 Based on JIS B7503; JIS B7524 For determining error of indicated size Comparison with ULMS / calibrated gauge blocks	Up to 25mm /0.001mm	3 µm	Laboratory
		Up to 25mm /0.01mm	6.5 µm	
Depth Gauge (Dial/Digital/ Vernier)	GTS-WP-29 Based on BS EN ISO 13385- 2 For determining error of indicated depth Comparison with HMS / gauge blocks	Up to 450 mm/0.02 mm	35 µm	Laboratory/ Customer Premises (Std. room/ Metrology)
		Up to 450 mm/0.01 mm	25 µm	
Height Gauge (Digital/Dial/ Analogue)	GTS-WP-25 Based on BS EN ISO 13225/ BS-1643 For determining error of indicated vertical size Comparison with gauge blocks and HMS	Up to 600mm	30 µm	Laboratory/ Customer Premises (Std. room/ Metrology)
		Up to 1000mm	40 µm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Feeler Gauge	GTS-WP-56 Based on BS 957 For determining thickness Comparison method using calibrated digital micrometer	Up to 1mm	3.5 μ m	Laboratory/ Customer Premises (Std. room/ Metrology)
Radius Gauge	GTS-WP-81 Based on IS 5273-1969 For determining radius using Profile Projector	Up to 25mm	9 μ m	Laboratory
Thread / Screw Pitch Gauge (Metric/inch)	GTS-WP-126 Based on IS 4211 For determining pitch using Profile Projector	0.4 - 7 mm	6 μ m	Laboratory
		4 - 42 TPI	240 μ in	
Thread Plug gauges (Metric / Unified/BSP (or) G threads)	GTS-WP-70 Based on EURAMET cg-10 For determining Simple Pitch Diameter using ULMS Metric Threads For determining Simple Pitch Diameter using ULMS Inch - Unified / BSP	1mm Up to 100mm	4 μ m	Laboratory
		>100 Up to 200mm	5 μ m	
		1/16" up to 4"	160 μ in	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Thread Plug gauges (Metric / Unified/BSP (or) G threads)	For determining Simple Pitch Diameter using ULMS Inch - Unified / BSP	Above 4" and including 8"	200 µin	Laboratory
	GTS-WP-137 Based on EURAMET cg-10 For determining Simple Pitch Diameter using ULMS Metric Threads	3mm up to 14mm	3 µm	
		>14mm up to and including 100 mm	4 µm	
	For determining Simple Pitch Diameter using ULMS Inch – Unified /BSP	1/8" up to 1/2"	120 µin	
		1/2"< and including 4"	160 µin	
Thread Plug/Ring gauge – Taper (NPT/BSPT)	GTS-WP-173 Based on JIS B 0262 & EURAMET cg-10 For determining Simple Pitch Diameter using ULMS	1/8" up to 1/2"	140 µin	Laboratory
		1/2"< and including 4"	180 µin	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Setting / Plain Plug gauge (Metric/Inch)	GTS-WP-80 Based on ASME B89.1.5 & EURAMET cg-6 For determining diameter Comparison with ULMS scale / gauge block using ULMS	1 mm up to 50 mm	1 μ m	Laboratory
		>50 mm up to 100 mm	1.5 μ m	
		>100 mm up to 400 mm	5 μ m	
Setting / Plain Ring gauge (Metric/Inch)	GTS-WP-106 Based on BS EN ISO 4064 & EURAMET cg-6 For determining diameter Comparison with reference ring gauge using ULMS	1 to 14mm	1.3 μ m	
		14< to 100mm	1.5 μ m	
		100< to 200mm	3 μ m	
		200< to 300mm	4.5 μ m	
		300< to 400mm	5 μ m	
Height Measuring System (HMS)/ Digital Height Gauge L.C: 0.001mm and greater	GTS-WP-169 Based on BS EN ISO 13225/BS 1643 For determining error of indicated vertical size Comparison with gauge blocks	Up to 1000mm	1+(0.008*L) μ m L: mm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
1-D measuring Machine (Universal Length Measuring System (ULMS))	GTS-WP-139 For determining error of indicated size/displacement Mechanical comparison to gauge blocks	Up to 100mm (absolute) Up to 600mm (differential)	0.2+(0.006*L) μm L: mm	Laboratory
Profile Projector	GTS-WP-158 Based on JIS 7540 For determining error of indicated size/displacement/angular displacement Comparison to calibrated Glass scale and Cross wire chart	Up to 200mm (0-360)°	5 + (0.015) μm L: mm 0.14° (8 arc minutes)	Laboratory
Thread Measuring Cylinder	GTS-WP-170 Based on BS 3777 & BS 5590 For determining diameter Comparison with ULMS scale / reference gauge block using ULMS	Up to 6.35mm	1 μm	Laboratory

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Cylindrical Standards / Measuring Pins	GTS-WP-170 Based on IS-11103 For determining diameter Comparison with ULMS scale / reference gauge block using ULMS	Up to 12 mm	1 μ m	Laboratory
Steel Scale	GTS-WP-171 Based on OIML R035-1-e& BS 4372 Measurement of line spacing using profile projector	Up to 300 mm	0.050 mm	
Caliper Checker	GTS-WP-164 Based on Manufacturer Spec. For determining face spacing Comparison to gauge blocks using precise HMS	Up to 600mm	1 + (0.01*L) μ m L: mm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Depth Micro Checker	GTS-WP-165 and 166 Based on Manufacturer Spec. For determining face spacing Comparison to gauge blocks using precise HMS	Up to 300mm	4.5 μm	Laboratory
Inside Micro checker	GTS-WP-166 Based on Manufacturer Spec. For determining face spacing Comparison to gauge blocks using precise HMS	Up to 300mm	4.5 μm	Laboratory
		>300mm Up to 600mm	7 μm	
Dial Calibration Tester	GTS-WP-167 Based on Manufactured Spec. For determining error of indicated displacement Mechanical comparison to gauge blocks using precise HMS or ULMS	Up to 25mm	2 μm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Dimensional Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Test Sieves	GTS-WP-43 Based on ISO 3310-1 For determining aperture size Using Profile Projector	50µm up to 4.3mm	8 µm	Laboratory
	GTS-WP-43 Based on ISO 3310-1 For determining aperture size using digital caliper	5mm up to 125mm	32 µm	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Voltage	Direct Method using Fluke 9100 <i>U = Measured voltage value</i>	000.001 mV to 010.000 Mv		Laboratory/ Customer Premises
		10 Hz to 3 kHz	$0.46 \times 10^3 U + 0.44 \text{ mV}$	
		>3 kHz to 10 kHz	$0.46 \times 10^3 U + 0.59 \text{ mV}$	
		>10 kHz to 30 kHz	$0.69 \times 10^3 U + 1.1 \text{ mV}$	
		>30 kHz to 50 kHz	$1.0 \times 10^3 U + 2.2 \text{ mV}$	
		>50 kHz to 100 kHz	$2.3 \times 10^3 U + 5.9 \text{ mV}$	
		010.001 mV to 032.000 mV		
		10 Hz to 3 kHz	$0.46 \times 10^3 U + 0.11 \text{ mV}$	
		>3 kHz to - 10 kHz	$0.4 \times 10^3 U + 0.15 \text{ mV}$	
		>10 kHz to 30 kHz	$0.70 \times 10^3 U + 0.28 \text{ mV}$	
		>30 kHz to 50 kHz	$1.0 \times 10^3 U + 0.56 \text{ mV}$	
		>50 kHz to 100 kHz	$2.3 \times 10^3 U + 1.5 \text{ mV}$	
		032.001 mV to 320.000 mV		
		10 Hz to 3 kHz	$0.47 \times 10^3 U + 22 \mu\text{V}$	
		>3 kHz to 10 kHz	$0.47 \times 10^3 U + 29 \mu\text{V}$	
		>10 kHz to 30 kHz	$0.70 \times 10^3 U + 56 \mu\text{V}$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Electrical Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Voltage	Direct Method using Fluke 9100 <i>U = Measured voltage value</i>	032.001 mV to 320.000 mV		Laboratory/ Customer Premises
		>30 kHz to 50 kHz	$1.1 \times 10^3 U + 0.11 \text{ mV}$	
		>50 kHz to 100 kHz	$2.3 \times 10^3 U + 0.30 \text{ mV}$	
		0.32001 V to 3.20000 V		
		10 Hz to 3 kHz	$0.48 \times 10^3 U + 0.22 \text{ mV}$	
		>3 kHz to 10 kHz	$0.47 \times 10^3 U + 0.29 \text{ mV}$	
		>10 kHz to 30 kHz	$0.70 \times 10^3 U + 0.55 \text{ mV}$	
		>30 kHz to 50 kHz	$1.1 \times 10^3 U + 1.1 \text{ mV}$	
		>50 kHz to 100 kHz	$2.3 \times 10^3 U + 3.0 \text{ mV}$	
		3.2001 V to 32.0000 V		
		10 Hz to 3 kHz	$0.48 \times 10^3 U + 2.2 \text{ mV}$	
		>3 kHz to 10 kHz	$0.71 \times 10^3 U + 2.9 \text{ mV}$	
		>10 kHz to 30 kHz	$0.93 \times 10^3 U + 5.5 \text{ mV}$	
		>30 kHz to 50 kHz	$1.7 \times 10^3 U + 11 \text{ mV}$	
		>50 kHz to 100 kHz	$4.1 \times 10^3 U + 37 \text{ mV}$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Electrical Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Voltage	Direct Method using Fluke 9100 <i>U = Measured voltage value</i>	032.001 V to 105.000 V		Laboratory/ Customer Premises
		10 Hz to 3 kHz	$0.47 \times 10^3 U + 7.3 \text{ mV}$	
		>3 kHz to 10 kHz	$0.70 \times 10^3 U + 9.7 \text{ mV}$	
		>10 kHz to 30 kHz	$0.93 \times 10^3 U + 18 \text{ mV}$	
		>30 kHz to 50 kHz	$1.7 \times 10^3 U + 36 \text{ mV}$	
		>50 kHz to 100 kHz	$4.1 \times 10^3 U + 0.12 \text{ V}$	
		105.001 V to 320.000 V		
		40 Hz to 100 Hz	$0.6 \times 10^3 U + 22 \text{ mV}$	
		>100 Hz to 1 kHz	$0.6 \times 10^3 U + 22 \text{ mV}$	
		>1 kHz to 3 kHz	$0.94 \times 10^3 U + 22 \text{ mV}$	
		>3 kHz to 10 kHz	$0.94 \times 10^3 U + 37 \text{ mV}$	
		>20 kHz to 30 kHz	$1.7 \times 10^3 U + 74 \text{ mV}$	
		0320.01 V to 0800.00 V		
		40 Hz to 100 Hz	$0.59 \times 10^3 U + 73 \text{ mV}$	
		>100 Hz to 1 kHz	$0.59 \times 10^3 U + 73 \text{ mV}$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Voltage	Direct Method using Fluke 9100 <i>U = Measured voltage value</i>	0320.01 V to 0800.00 V		Laboratory/ Customer Premises
		>1 kHz to 3 kHz	$0.93 \times 10^3 U + 73 \text{ mV}$	
		>3 kHz to 10 kHz	$0.93 \times 10^3 U + 0.12 \text{ V}$	
		0800.01 V to 1050.00 V		
		40 Hz to 100 Hz	$0.59 \times 10^3 U + 0.15 \text{ V}$	
		>100 Hz to 1 kHz	$0.59 \times 10^3 U + 0.15 \text{ V}$	
		>1 kHz to 3 kHz	$0.93 \times 10^3 U + 0.15 \text{ V}$	
		>3 kHz to 10 kHz	$0.93 \times 10^3 U + 0.24 \text{ V}$	
	Direct Method using Fluke 5522A <i>U = Measured voltage value</i>	1.0 mV to 32.999 mV		
		10 Hz to 45 Hz	$0.62 \times 10^{-3} U + 4.8 \mu\text{V}$	
		>45 Hz to 10 kHz	$0.13 \times 10^{-3} U + 4.8 \mu\text{V}$	
		>10 kHz to 20 kHz	$0.17 \times 10^{-3} U + 4.8 \mu\text{V}$	
		>20 kHz to 50 kHz	$0.78 \times 10^{-3} U + 4.8 \mu\text{V}$	
		>50 kHz to 100 kHz	$2.7 \times 10^{-3} U + 9.4 \mu\text{V}$	
	>100 kHz to 500 kHz	$6.3 \times 10^{-3} U + 39 \mu\text{V}$		

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Electrical Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Voltage	Direct Method using Fluke 5522A <i>U = Measured voltage value</i>	33 mV to 329.999 mV		Laboratory/ Customer Premises
		10 Hz to 45 Hz	$0.24 \times 10^{-3} U + 6.1 \mu\text{V}$	
		>45 Hz to 10 kHz	$0.13 \times 10^{-3} U + 6.0 \mu\text{V}$	
		>10 kHz to 20 kHz	$0.14 \times 10^{-3} U + 6.0 \mu\text{V}$	
		>20 kHz to 50 kHz	$0.28 \times 10^{-3} U + 6.0 \mu\text{V}$	
		>50 kHz to 100 kHz	$0.63 \times 10^{-3} U + 25 \mu\text{V}$	
		>100 kHz to 500 kHz	$1.7 \times 10^{-3} U + 53 \mu\text{V}$	
		0.33 V to 3.29999 V		
		10 Hz to 45 Hz	$0.48 \times 10^{-3} U + 80 \mu\text{V}$	
		>45 Hz to 10 kHz	$0.13 \times 10^{-3} U + 45 \mu\text{V}$	
		>10 kHz to 20 kHz	$0.16 \times 10^{-3} U + 45 \mu\text{V}$	
		>20 kHz to 50 kHz	$0.24 \times 10^{-3} U + 38 \mu\text{V}$	
		>50 kHz to 100 kHz	$0.55 \times 10^{-3} U + 97 \mu\text{V}$	
		>100 kHz to 500 kHz	$1.9 \times 10^{-3} U + 0.46 \text{ mV}$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Voltage	Direct Method using Fluke 5522A <i>U = Measured voltage value</i>	3.3 V to 32.9999 V		Laboratory/ Customer Premises
		10 Hz to 45 Hz	$0.48 \times 10^{-3} U + 1.0 \text{ mV}$	
		>45 Hz to 10 kHz	$0.13 \times 10^{-3} U + 0.45 \text{ mV}$	
		>10 kHz to 20 kHz	$0.20 \times 10^{-3} U + 0.46 \text{ mV}$	
		>20 kHz to 50 kHz	$0.28 \times 10^{-3} U + 0.46 \text{ mV}$	
		>50 kHz to 100 kHz	$0.71 \times 10^{-3} U + 1.2 \text{ mV}$	
		33 V to 329.9999 V		
		45 Hz to 1 kHz	$0.16 \times 10^{-3} U + 1.6 \text{ mV}$	
		>1 kHz to 10 kHz	$0.17 \times 10^{-3} U + 4.5 \text{ mV}$	
		>10 kHz to 20 kHz	$0.21 \times 10^{-3} U + 4.6 \text{ mV}$	
		>20 kHz to 50 kHz	$0.26 \times 10^{-3} U + 4.4 \text{ mV}$	
		>50 kHz to 100 kHz	$1.6 \times 10^{-3} U + 39 \text{ mV}$	
		330 V to 1020 V		
		45 Hz to 1 kHz	$0.24 \times 10^{-3} U + 8.5 \text{ mV}$	
		>1 kHz to 5 kHz	$0.21 \times 10^{-3} U + 8.5 \text{ mV}$	
		>5 kHz to 10 kHz	$0.26 \times 10^{-3} U + 8.0 \text{ mV}$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
DC Voltage	Direct Method using Fluke 9100 <i>U = Measured Voltage value</i>	0.001 mV to 320.000 mV	$14 \times 10^6 U + 1,7 \mu V$	Laboratory/ Customer Premises
		0.32001 V to 3.20000 V	$9,0 \times 10^6 U + 1,7 \mu V$	
		3.2001 V to 32.0000 V	$10 \times 10^6 U + 17 \mu V$	
		32.001 V to 320.000 V	$15 \times 10^6 U + 0,13 mV$	
		320.01 V to 1050.00 V	$15 \times 10^6 U + 1,3 mV$	
	Direct Method using Fluke 5522A <i>U = Measured Voltage value</i>	0 V to 329.9999 mV	$56 \times 10^{-6} U + 2 \mu V$	
		0.33 V to 3.299999 V	$58 \times 10^{-6} U + 0,3 \mu V$	
		3.3 V to 32.99999 V	$59 \times 10^{-6} U + 3,5 \mu V$	
		33 V to 329.9999 V	$60 \times 10^{-6} U + 0,035 mV$	
		330 to 1020.000 V	$60 \times 10^{-6} U + 0,36 mV$	
DC Current	Direct Method using Fluke 9100 <i>I = Measured Current value</i>	0.001 μA to 320.000 μA	$0,17 \times 10^3 / + 0,013 \mu A$	
		0.32001 mA to 3.20000 mA	$0,18 \times 10^3 / + 0,094 \mu A$	
		3.2001 mA to 32.0000 mA	$0,18 \times 10^3 / + 1,0 \mu A$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
DC Current	Direct Method using Fluke 9100 <i>I = Measured Current value</i>	32.001 mA to 320.000 mA	$0.20 \times 10^{-3} / + 11 \mu\text{A}$	Laboratory/ Customer Premises
		0.32001 A to 3.20000 A	$0.70 \times 10^{-3} / + 0.14 \text{ mA}$	
		3.2001 A to 10.5000 A	$0.64 \times 10^{-3} / + 1.1 \text{ mA}$	
		10-turn current coil		
		3.2001 A to 32.0000 A	$2.4 \times 10^{-3} / + 0.51 \text{ mA}$	
		32.001 A to 105.000 A	$2.4 \times 10^{-3} / + 3.8 \text{ mA}$	
		50-turn current coil		
		16.001 A to 160.000 A	$2.4 \times 10^{-3} / + 2.5 \text{ mA}$	
		160.01 A to 525.00 A	$2.4 \times 10^{-3} / + 19 \text{ mA}$	
	525.01 A to 1000.00 A	$2.4 \times 10^{-3} / + 0.104 \text{ A}$		
	Direct Method using Fluke 5522A <i>I = Measured Current value</i>	0 μA to 329.999 μA	$0.12 \times 10^{-3} / + 0.011 \mu\text{A}$	
		0.33 mA to 3.29999 mA	$0.08 \times 10^{-3} / + 0.04 \mu\text{A}$	
		3.3 mA to 32.99999 mA	$0.08 \times 10^{-3} / + 0.21 \mu\text{A}$	
		33 mA to 329.999 mA	$0.16 \times 10^{-3} / + 4.2 \mu\text{A}$	
		0.33 A to 1.09999 A	$0.16 \times 10^{-3} / + 0.031 \text{ mA}$	
1.1 A to 2.99999 A		$0.30 \times 10^{-3} / + 0.031 \text{ mA}$		
3 A to 10.9999 A		$0.41 \times 10^{-3} / + 0.38 \text{ mA}$		
11 A to 20.5 A		$0.85 \times 10^{-3} / + 0.54 \text{ mA}$		

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location	
Calibration of instruments					
DC Current	Direct Method using Fluke 5522A <i>I = Measured Current value</i>	50 turn coil		Laboratory/ Customer Premises	
		0.2 A to 0.33 A	$4.0 \times 10^{-3} / +16 \text{ mA}$		
		>0.33 A to 2.9999 A	$4.0 \times 10^{-3} / +0.11 \text{ A}$		
		3 A to 20.5 A	$4.0 \times 10^{-3} / +0.39 \text{ A}$		
AC Current	Direct Method using Fluke 9100 <i>I = Measured Current value</i>	0.001 μA to 320.000 μA			
		10 Hz to 3 kHz	$0.83 \times 10^{-3} / + 0.35 \mu\text{A}$		
		>3 kHz to 10 kHz	$1.2 \times 10^{-3} / + 0.69 \mu\text{A}$		
		>10 kHz to 20 kHz	$2.3 \times 10^{-3} / + 2.3 \mu\text{A}$		
		>20 kHz to 30 kHz	$2.9 \times 10^{-3} / + 3.5 \mu\text{A}$		
		0.32001 mA to 3.20000 mA			
		10 Hz to 3 kHz	$0.85 \times 10^{-3} / + 0.34 \mu\text{A}$		
		>3 kHz to 10 kHz	$1.2 \times 10^{-3} / + 0.68 \mu\text{A}$		
		>10 kHz to 20 kHz	$2.4 \times 10^{-3} / + 2.3 \mu\text{A}$		
		>20 kHz to 30 kHz	$2.9 \times 10^{-3} / + 3.5 \mu\text{A}$		

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location	
Calibration of instruments					
AC Current	Direct Method using Fluke 9100 <i>I = Measured Current value</i>	10 Hz to 3 kHz	$0.85 \times 10^{-3} / + 3.6 \mu\text{A}$	Laboratory/ Customer Premises	
		>3 kHz to 10 kHz	$1.2 \times 10^{-3} / + 7.3 \mu\text{A}$		
		>10 kHz to 20 kHz	$2.4 \times 10^{-3} / + 15 \mu\text{A}$		
		>20 kHz to 30 kHz	$2.9 \times 10^{-3} / + 26 \mu\text{A}$		
		32.001 mA to 320.000 mA			
		10 Hz to 3 kHz	$0.9 \times 10^{-3} / + 36 \mu\text{A}$		
		>3 kHz to 10 kHz	$1.2 \times 10^{-3} / + 54 \mu\text{A}$		
		>10 kHz to 20 kHz	$2.4 \times 10^{-3} / + 74 \mu\text{A}$		
		>20 kHz to 30 kHz	$2.9 \times 10^{-3} / + 0.11 \text{ mA}$		
		0.32001 A to 3.20000 A			
		10 Hz to 3 kHz	$1.2 \times 10^{-3} / + 0.55 \mu\text{A}$		
		>3 kHz to 10 kHz	$2.9 \times 10^{-3} / + 3 \text{ mA}$		
		3.2001 A to 10.5000 A			
		10 Hz to 3 kHz	$2.3 \times 10^{-3} / + 3.5 \text{ mA}$		
		>3 kHz to 10 kHz	$5.8 \times 10^{-3} / + 12 \text{ mA}$		

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Calibration of instruments

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
AC Current	Direct Method using Fluke 9100 <i>I = Measured Current value</i>	10-turn current coil		Laboratory/ Customer Premises
		3.2001 A to 32.0000 A		
		10 Hz to 100 Hz	$3.3 \times 10^{-3} / + 5.2 \text{ mA}$	
		>100 Hz to 440 Hz	$9.3 \times 10^{-3} / + 31 \text{ mA}$	
		32.001 A to 200.000 A		
		10 Hz to 100 Hz	$3.3 \times 10^{-3} / + 91 \text{ mA}$	
		>100 Hz to 440 Hz	$8.1 \times 10^{-3} / + 0.28 \text{ A}$	
		50-turn current coil		
		16.001 A to 160.000 A		
		10 Hz to 100 Hz	$3.3 \times 10^{-3} / + 27 \text{ mA}$	
	160.01 A to 1000.00 A			
	10 Hz to 100 Hz	$3.3 \times 10^{-3} / + 0.45 \text{ A}$		
	Direct Method using Fluke 5522A <i>I = Measured Current value</i>	29 μA to 329.99 μA		
		10 Hz to 20 Hz	$1.6 \times 10^{-3} / +0.10 \mu\text{A}$	
		>20 Hz to 45 Hz	$1.2 \times 10^{-3} / +0.10 \mu\text{A}$	
		>45 Hz to 1 kHz	$0.97 \times 10^{-3} / +0.10 \mu\text{A}$	
>1 kHz to 5 kHz		$2.3 \times 10^{-3} / +0.12 \mu\text{A}$		

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Electrical Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Current	Direct Method using Fluke 5522A <i>I = Measured Current value</i>	29 μA to 329.99 μA		Laboratory/ Customer Premises
		>5 kHz to 10 kHz	$6.2 \times 10^{-3} / +0.16 \mu\text{A}$	
		>10 kHz to 30 kHz	$12 \times 10^{-3} / +0.31 \mu\text{A}$	
		0.33 mA to 3.29999 mA		
		10 Hz to 20 Hz	$1.6 \times 10^{-3} / +0.12 \mu\text{A}$	
		>20 Hz to 45 Hz	$0.97 \times 10^{-3} / +0.12 \mu\text{A}$	
		>45 Hz to 1 kHz	$0.78 \times 10^{-3} / +0.12 \mu\text{A}$	
		>1 kHz to 5 kHz	$1.55 \times 10^{-3} / +0.16 \mu\text{A}$	
		>5 kHz to 10 kHz	$3.9 \times 10^{-3} / +0.23 \mu\text{A}$	
		>10 kHz to 30 kHz	$7.8 \times 10^{-3} / +0.46 \mu\text{A}$	
		3.3 mA to 32.9999 mA		
		10 Hz to 20 Hz	$1.4 \times 10^{-3} / +1.6 \mu\text{A}$	
		>20 Hz to 45 Hz	$0.71 \times 10^{-3} / +1.5 \mu\text{A}$	
		>45 Hz to 1 kHz	$0.35 \times 10^{-3} / +1.5 \mu\text{A}$	
		>1 kHz to 5 kHz	$0.69 \times 10^{-3} / +1.5 \mu\text{A}$	
		>5 kHz to 10 kHz	$1.6 \times 10^{-3} / +1.5 \mu\text{A}$	
>10 kHz to 30 kHz	$3.1 \times 10^{-3} / +1.5 \mu\text{A}$			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Electrical Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Current	Direct Method using Fluke 5522A <i>I = Measured Current value</i>	33 mA to 329.999 mA		Laboratory/ Customer Premises
		10 Hz to 20 Hz	$1.4 \times 10^{-3} / +16 \mu\text{A}$	
		>20 Hz to 45 Hz	$0.70 \times 10^{-3} / +16 \mu\text{A}$	
		>45 Hz to 1 kHz	$0.32 \times 10^{-3} / +15 \mu\text{A}$	
		>1 kHz to 5 kHz	$0.78 \times 10^{-3} / +39 \mu\text{A}$	
		>5 kHz to 10 kHz	$1.6 \times 10^{-3} / +78 \mu\text{A}$	
		>10 kHz to 30 kHz	$3.1 \times 10^{-3} / +0.16 \text{ mA}$	
		0.33 A to 1.09999 A		
		10 Hz to 45 Hz	$1.4 \times 10^{-3} / +76 \mu\text{A}$	
		>45 Hz to 1 kHz	$0.41 \times 10^{-3} / +76 \mu\text{A}$	
		>1 kHz to 5 kHz	$4.7 \times 10^{-3} / +0.77 \text{ mA}$	
		>5 kHz to 10 kHz	$19 \times 10^{-3} / +3.9 \text{ mA}$	
		1.11 A to 2.99999 A		
		10 Hz to 45 Hz	$1.4 \times 10^{-3} / +77 \mu\text{A}$	
		>45 Hz to 1 kHz	$0.48 \times 10^{-3} / +76 \mu\text{A}$	
		>1 kHz to 5 kHz	$4.7 \times 10^{-3} / +0.77 \text{ mA}$	
		>5 kHz to 10 kHz	$19 \times 10^{-3} / +3.9 \text{ mA}$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
AC Current	Direct Method using Fluke 5522A <i>I = Measured Current value</i>	3 A to 10.9999 A		Laboratory/ Customer Premises
		45Hz to 100 Hz	$0.48 \times 10^{-3} / +1.5 \text{ mA}$	
		>100 Hz to 1 kHz	$0.79 \times 10^{-3} / +1.5 \text{ mA}$	
		>1 kHz to 5 kHz	$23 \times 10^{-3} / +1.6 \text{ mA}$	
		11 A to 20.5 A		
		45 Hz to 100 Hz	$0.95 \times 10^{-3} / +3.8 \text{ mA}$	
		>100 Hz to 1 kHz	$1.2 \times 10^{-3} / +3.8 \text{ mA}$	
		>1 kHz to 5 kHz	$23 \times 10^{-3} / +3.9 \text{ mA}$	
Resistance	Direct Method using Fluke 9100 <i>R = Measured Resistance value</i>	0.0001 Ω to 40.0000 Ω	$0.33 \times 10^{-3} R + 12 \text{ m}\Omega$	
		40.001 Ω to 400.000 Ω	$0.23 \times 10^{-3} R + 23 \text{ m}\Omega$	
		0.40001 k Ω to 4.00000 k Ω	$0.17 \times 10^{-3} R + 93 \text{ m}\Omega$	
		4.0001 k Ω to 40.0000 k Ω	$0.17 \times 10^{-3} R + 0.93 \Omega$	
		40.001 k Ω to 400.000 k Ω	$0.21 \times 10^{-3} R + 9.3 \Omega$	
		0.40001 M Ω to 4.00000 M Ω	$0.23 \times 10^{-3} R + 0.12 \text{ k}\Omega$	
		4.0001 M Ω to 40.0000 M Ω	$0.59 \times 10^{-3} R + 2.3 \text{ k}\Omega$	
		40.001 M Ω to 400.000 M Ω	$0.71 \times 10^{-3} R + 46 \text{ k}\Omega$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
Resistance	Direct Method using Fluke 5522A <i>R = Measured Resistance value</i>	0 Ω to 10.9999 Ω	$31 \times 10^{-6} R + 7.8 \text{ m}\Omega$	Laboratory/ Customer Premises
		11 Ω to 32.9999 Ω	$23 \times 10^{-6} R + 12 \text{ m}\Omega$	
		33 Ω to 109.9999 Ω	$22 \times 10^{-6} R + 12 \text{ m}\Omega$	
		110 Ω to 329.9999 Ω	$22 \times 10^{-6} R + 16 \text{ m}\Omega$	
		330 Ω to 1.099999 kΩ	$22 \times 10^{-6} R + 15 \text{ m}\Omega$	
		1.1 kΩ to 3.299999 kΩ	$22 \times 10^{-6} R + 0.15 \Omega$	
		3.3 kΩ to 10.99999 kΩ	$22 \times 10^{-6} R + 0.077 \Omega$	
		11 kΩ to 32.99999 kΩ	$22 \times 10^{-6} R + 0.77 \Omega$	
		33 kΩ to 109.9999 kΩ	$22 \times 10^{-6} R + 0.77 \Omega$	
		110 kΩ to 329.9999 kΩ	$25 \times 10^{-6} R + 7.7 \Omega$	
		330 kΩ to 1.099999 MΩ	$26 \times 10^{-6} R + 7.7 \Omega$	
		1.1 MΩ to 3.299999 MΩ	$48 \times 10^{-6} R + 0.12 \text{ k}\Omega$	
		3.3 MΩ to 10.99999 MΩ	$0.10 \times 10^{-3} R + 0.19 \text{ k}\Omega$	
		11 MΩ to 32.99999 MΩ	$0.21 \times 10^{-3} R + 1.9 \text{ k}\Omega$	
		33 MΩ to 109.9999 MΩ	$0.44 \times 10^{-3} R + 2.1 \text{ k}\Omega$	
110 MΩ to 329.9999 MΩ	$2.3 \times 10^{-3} R + 0.077 \text{ M}\Omega$			
330 MΩ to 1100 MΩ	$12 \times 10^{-3} R + 0.39 \text{ M}\Omega$			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
Capacitance	Direct Method using Fluke 9100	0.5000 nF to 4.0000 nF	$3.5 \times 10^{-3} C + 18 \text{ pF}$	Laboratory/ Customer Premises
		4.0001 nF to 40.000 nF	$3.5 \times 10^{-3} C + 35 \text{ pF}$	
		40.001 nF to 400.00 nF	$3.5 \times 10^{-3} C + 0.18 \text{ nF}$	
		400.01 nF to 4.0000 μF	$4.7 \times 10^{-3} C + 1.9 \text{ nF}$	
		4.0001 μF to 40.000 μF	$5.8 \times 10^{-3} C + 19 \text{ nF}$	
		40.001 μF to 400.00 μF	$5.8 \times 10^{-3} C + 0.19 \mu\text{F}$	
		400.01 μF to 4.0000 mF	$5.8 \times 10^{-3} C + 1.8 \mu\text{F}$	
		4.0001 mF to 40.000 mF	$12 \times 10^{-3} C + 69 \mu\text{F}$	
	Direct Method using Fluke 5522A <i>C = Measured Capacitance value</i>	220 pF to 399.9 pF	$8.6 \times 10^{-3} C + 7.2 \text{ pF}$	
		0.4 nF to 1.0999 nF	$4.5 \times 10^{-3} C + 7.6 \text{ pF}$	
		1.1 nF to 3.2999 nF	$4.1 \times 10^{-3} C + 7.6 \text{ pF}$	
		3.3 nF to 10.9999 nF	$2.1 \times 10^{-3} C + 7.6 \text{ pF}$	
		11 nF to 32.9999 nF	$2.0 \times 10^{-3} C + 77 \text{ pF}$	
		33 nF to 109.999 nF	$2.1 \times 10^{-3} C + 76 \text{ pF}$	
		110 nF to 329.999 nF	$2.1 \times 10^{-3} C + 0.23 \text{ nF}$	
0.33 μF to 1.09999 μF	$2.1 \times 10^{-3} C + 0.76 \text{ nF}$			
1.1 μF to 3.29999 μF	$2.1 \times 10^{-3} C + 2.3 \text{ nF}$			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of instruments				
Capacitance	Direct Method using Fluke 5522A <i>C = Measured Capacitance value</i>	3.3 µF to 10.9999 µF	$2.1 \times 10^{-3} C + 7.6 \text{ nF}$	Laboratory/ Customer Premises
		11 µF to 32.9999 µF	$3.2 \times 10^{-3} C + 23 \text{ nF}$	
		33 µF to 109.999 µF	$3.7 \times 10^{-3} C + 75 \text{ nF}$	
		110 µF to 329.999 µF	$3.7 \times 10^{-3} C + 0.22 \text{ µF}$	
		0.33 mF - 1.09999 mF	$5.4 \times 10^{-3} C + 0.75 \text{ µF}$	
		1.1 mF to 3.29999 mF	$5.4 \times 10^{-3} C + 2.2 \text{ µF}$	
		3.3 mF to 10.9999 mF	$5.4 \times 10^{-3} C + 7.5 \text{ µF}$	
		11 mF to 32.9999 mF	$8.8 \times 10^{-3} C + 23 \text{ µF}$	
		33 mF to 110 mF	$13 \times 10^{-3} C + 77 \text{ µF}$	
Frequency	Direct Method using Fluke 9100 Direct Method using Fluke 5522A <i>f = Measured Frequency value</i>	0.5 Hz to 10.0 MHz	$29 \times 10^{-6} f$	
		0.01 Hz to 119.99 Hz	$1.9 \times 10^{-6} f + 12 \text{ µHz}$	
		120 Hz to 1199.9 Hz	$2.0 \times 10^{-6} f + 32 \text{ µHz}$	
		1.200 kHz to 11.999 kHz	$2.0 \times 10^{-6} f + 0.29 \text{ mHz}$	
		12.00 kHz to 119.99 kHz	$2.0 \times 10^{-6} f + 2.9 \text{ mHz}$	
		120.00 kHz to 1199.9 kHz	$2.0 \times 10^{-6} f + 29 \text{ mHz}$	
		1.200 MHz to 2.000 MHz	$1.9 \times 10^{-6} f + 0.42 \text{ Hz}$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
RTD - simulation	Simulation Method using Fluke 5522A	-200 °C to -80 °C RTD-pt385, 100Ω	0.04 °C	Laboratory/ Customer Premises
		> -80 °C to 0.003 °C RTD-pt385, 100Ω	0.04 °C	
		0.03 °C to 100 °C RTD-pt385, 100Ω	0.06 °C	
		>100 °C to 300 °C RTD-pt385, 100Ω	0.07 °C	
		>300 °C to 400 °C RTD-pt385, 100Ω	0.08 °C	
		>400 °C to 630 °C RTD-pt385, 100Ω	0.09 °C	
		>630 °C to 800 °C RTD-pt385, 100Ω	0.18 °C	
		-200 °C to -80 °C RTD-pt3926, 100Ω	0.04 °C	
		>-80 °C to 0.003 °C RTD-pt3926, 100Ω	0.06 °C	
		0.03 °C to 100 °C / RTD-pt3926, 100Ω	0.07 °C	
		>100 °C to 300 °C RTD-pt3926, 100Ω	0.08 °C	
		>300 °C to 400 °C RTD-pt3926, 100Ω	0.09 °C	
		>400 °C to 630 °C RTD-pt3926, 100Ω	0.18 °C	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
RTD - simulation	Simulation Method using Fluke 5522A	-200 °C to -190 °C RTD-pt3916, 100Ω	0.19 °C	Laboratory/ Customer Premises
		>-190 °C to -80 °C RTD-pt3916, 100Ω	0.03 °C	
		>-80 °C to 0.003 °C RTD-pt3916, 100Ω	0.04 °C	
		0.03 °C to 100 °C RTD-pt3916, 100Ω	0.05 °C	
		>100 °C to 260 °C RTD-pt3916, 100Ω	0.06 °C	
		>260 °C to 300 °C RTD-pt3916, 100Ω	0.06 °C	
		>300 °C to 400 °C RTD-pt3916, 100Ω	0.07 °C	
		>400 °C to 600 °C RTD-pt3916, 100Ω	0.08 °C	
		>600 °C to 630 °C RTD-pt3916, 100Ω	0.18 °C	
		-200 °C to -80 °C RTD-pt385, 200Ω	0.03 °C	
		>-80 °C to 0.003 °C RTD-pt385, 200Ω	0.03 °C	
		0.03 °C to 100 °C RTD-pt385, 200Ω	0.03 °C	
		>100 °C to 260 °C RTD-pt385, 200Ω	0.04 °C	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
RTD - simulation	Simulation Method using Fluke 5522A	>260 °C to 300 °C RTD-pt385, 200Ω	0.09 °C	Laboratory/ Customer Premises
		>300 °C to 400 °C RTD-pt385, 200Ω	0.10 °C	
		>400 °C to 600 °C RTD-pt385, 200Ω	0.11 °C	
		>600 °C to 630 °C RTD-pt385, 200Ω	0.12 °C	
		-200 °C to -80 °C RTD-pt385, 500Ω	0.03 °C	
		>-80 °C to 0.003 °C RTD-pt385, 500Ω	0.04 °C	
		0.03 °C to 100 °C RTD-pt385, 500Ω	0.04 °C	
		>100 °C to 260 °C RTD-pt385, 500Ω	0.05 °C	
		>260 °C to 300 °C RTD-pt385, 500Ω	0.06 °C	
		>300 °C to 400 °C RTD-pt385, 500Ω	0.06 °C	
		>400 °C to 600 °C RTD-pt385, 500Ω	0.07 °C	
		>600 °C to 630 °C RTD-pt385, 500Ω	0.09 °C	
-200 °C to -80 °C RTD-pt385, 1000Ω	0.03 °C			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
RTD - simulation	Simulation Method using Fluke 5522A	> -80 °C to 0.003 °C RTD-pt385, 1000Ω	0.03 °C	Laboratory/ Customer Premises
		0.03 °C to 100 °C RTD-pt385, 1000Ω	0.03 °C	
		>100 °C to 260 °C RTD-pt385, 1000Ω	0.04 °C	
		>260 °C to 300 °C RTD-pt385, 1000Ω	0.05 °C	
		>300 °C to 400 °C RTD-pt385, 1000Ω	0.06 °C	
		>400 °C to 600 °C RTD-pt385, 1000Ω	0.06 °C	
		>600 °C to 630 °C RTD-pt385, 1000Ω	0.18 °C	
		-80 °C to 0.003 °C RTD-pt385, 120Ω (Ni120)	0.03 °C	
		0.03 °C to 100 °C RTD-pt385, 120Ω (Ni120)	0.03 °C	
		>100 °C to 260 °C RTD-pt385, 120Ω (Ni120)	0.04 °C	
		-100 °C to 260 °C RTD-Cu427, 10Ω	0.23 °C	
		600 °C to 800 °C	0.34 °C	
		>800 °C to 1000 °C	0.26 °C	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Thermocouple B	Simulation Method using Fluke 5522A	600 °C to 800 °C	0.34 °C	Laboratory/ Customer Premises
		>800 °C to 1000 °C	0.26 °C	
		>1000 °C to 1550 °C	0.23 °C	
		>1550 °C to 1820 °C	0.26 °C	
Thermocouple C	Simulation Method using Fluke 5522A	0.01 °C to 150 °C	0.23 °C	
		>150 °C to 650 °C	0.20 °C	
		>650 °C to 1000 °C	0.24 °C	
		>1000 °C to 1800°C	0.39 °C	
		>1800 °C to 2316°C	0.65 °C	
Thermocouple E	Simulation Method using Fluke 5522A	-250 °C to -100 °C	0.39 °C	
		>-100 °C to -25 °C	0.12 °C	
		>-25 °C to 350 °C	0.11 °C	
		>350 °C to 650°C	0.12 °C	
		>650 °C to 1000°C	0.16 °C	
Thermocouple J	Simulation Method using Fluke 5522A	-210 °C to -100 °C	0.21 °C	
		>-100 °C to -35 °C	0.12 °C	
		>-30 °C to 150 °C	0.11 °C	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Thermocouple J	Simulation Method using Fluke 5522A	>150 °C to 760°C	0.13 °C	Laboratory/ Customer Premises
		>760 °C to 1200°C	0.18 °C	
Thermocouple K	Simulation Method using Fluke 5522A	-200 °C to -100 °C	0.26 °C	
		>-100 °C to -25 °C	0.14 °C	
		>-25 °C to 120 °C	0.12 °C	
		>120 °C to 1000 °C	0.20 °C	
		>1000 °C to 1372 °C	0.31 °C	
Thermocouple L	Simulation Method using Fluke 5522A	-200 °C to -100 °C	0.29 °C	
		>-100 °C to 800 °C	0.20 °C	
		>800°C to 900 °C	0.13 °C	
Thermocouple N	Simulation Method using Fluke 5522A	-200 °C to -100 °C	0.31 °C	
		>-100 °C to -25 °C	0.17 °C	
		>-25 °C to 120 °C	0.15 °C	
		>120 °C to 410 °C	0.14 °C	
		>410 °C to 1300 °C	0.21 °C	
Thermocouple R	Simulation Method using Fluke 5522A	0.01 °C to 250 °C	0.44 °C	
		>250 °C to 400 °C	0.27 °C	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Thermocouple R	Simulation Method using Fluke 5522A	>400 °C to 1000 °C	0.26 °C	Laboratory/ Customer Premises
		>1000 °C to 1767°C	0.31 °C	
Thermocouple S	Simulation Method using Fluke 5522A	0.01 °C to 250 °C	0.36 °C	
		>250 °C to 1000 °C	0.28 °C	
		>1000 °C to 1400 °C	0.29 °C	
		>1400 °C to 1767°C	0.36 °C	
Thermocouple T	Simulation Method using Fluke 5522A	-250 °C to -150 °C	0.49 °C	
		> -150 °C to 0.003 °C	0.19 °C	
		0.01 °C to 120 °C	0.12 °C	
		>120 °C to 400°C	0.11 °C	
Thermocouple U	Simulation Method using Fluke 5522A	-200 °C to 0.01 °C	0.43 °C	
		>0.01 °C to 600 °C	0.21 °C	
DC Power	Direct Method using Fluke 5522A with PQ Option	33 mV/0.33mA	$0.28 \times 10^{-3} P$	
		33 mV/329.99 mA	$0.20 \times 10^{-3} P$	
		1020 V/0.33mA	$0.29 \times 10^{-3} P$	
		1020 V/329.99 mA	$0.20 \times 10^{-3} P$	
		33 mV/0.33 A	$0.40 \times 10^{-3} P$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Electrical Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
DC Power	Direct Method using Fluke 5522A with PQ Option	33 mV/2.9999 A	$0.48 \times 10^{-3} P$	Laboratory/ Customer Premises
		1020 V/0.33 A	$0.40 \times 10^{-3} P$	
		1020 V/2.9999 A	$0.48 \times 10^{-3} P$	
		33 mV/3 A	$0.83 \times 10^{-3} P$	
		33 mV/20.5 A	$1.3 \times 10^{-3} P$	
		1020 V/3 A	$0.84 \times 10^{-3} P$	
		1020 V/20.5 A	$1.3 \times 10^{-3} P$	
AC Power	Direct Method using Fluke 5522A with PQ Option	45 Hz to 65 Hz PF=1		
		33 mV/3.3mA	$1.4 \times 10^{-3} P$	
		33 mV/8.999 mA	$1.0 \times 10^{-3} P$	
		33 mV/9 mA	$0.92 \times 10^{-3} P$	
		33 mV/32.999 mA	$0.78 \times 10^{-3} P$	
		33 mV/33 mA	$1.3 \times 10^{-3} P$	
		33 mV/89.99 mA	$0.99 \times 10^{-3} P$	
		33 mV/90 mA	$0.91 \times 10^{-3} P$	
		33 mV/329.99 mA	$0.77 \times 10^{-3} P$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
AC Power	Direct Method using Fluke 5522A with PQ Option	329.99 mV/3.3 mA	$1.3 \times 10^{-3} P$	Laboratory/ Customer Premises
		329.99 mV/8.999 mA	$0.94 \times 10^{-3} P$	
		329.999 mV/9 mA	$0.86 \times 10^{-3} P$	
		329.999 mV/32.999 mA	$0.71 \times 10^{-3} P$	
		329.999 mV/33 mA	$1.3 \times 10^{-3} P$	
		329.999 mV/89.99 mA	$0.93 \times 10^{-3} P$	
		329.999 mV/90 mA	$0.85 \times 10^{-3} P$	
		329.999 mV/329.99 mA	$0.69 \times 10^{-3} P$	
		330 mV/3.3 mA	$1.4 \times 10^{-3} P$	
		330 mV/8.999 mA	$0.95 \times 10^{-3} P$	
		330 mV/9 mA	$0.89 \times 10^{-3} P$	
		330 mV/32.999 mA	$0.75 \times 10^{-3} P$	
		330 mV/33 mA	$1.3 \times 10^{-3} P$	
		330 mV/89.99 mA	$0.94 \times 10^{-3} P$	
		330 mV/90 mA	$0.88 \times 10^{-3} P$	
		330 mV/329.99 mA	$0.73 \times 10^{-3} P$	
1020 V/3.3 mA	$1.3 \times 10^{-3} P$			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
AC Power	Direct Method using Fluke 5522A with PQ Option	1020 V/8.999 mA	$0.95 \times 10^{-3} P$	Laboratory/ Customer Premises
		1020 V/9 mA	$0.88 \times 10^{-3} P$	
		1020 V/32.999 mA	$0.73 \times 10^{-3} P$	
		1020 V/33 mA	$1.3 \times 10^{-3} P$	
		1020 V/89.99 mA	$0.93 \times 10^{-3} P$	
		1020 V/90 mA	$0.87 \times 10^{-3} P$	
		1020 V/329.99 mA	$0.72 \times 10^{-3} P$	
		33 mV/0.33 A	$1.1 \times 10^{-3} P$	
		33 mV/0.8999 A	$0.96 \times 10^{-3} P$	
		33 mV/0.9 A	$0.92 \times 10^{-3} P$	
		33 mV/2.1999 A	$0.95 \times 10^{-3} P$	
		33 mV/2.2A	$0.99 \times 10^{-3} P$	
		33 mV/4.4999 A	$1.4 \times 10^{-3} P$	
		33 mV/4.5A	$1.4 \times 10^{-3} P$	
		33 mV/20.5 A	$1.8 \times 10^{-3} P$	
329.999 mV/0.33 A	$1.1 \times 10^{-3} P$			
329.999 mV/0.8999 A	$0.89 \times 10^{-3} P$			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
AC Power	Direct Method using Fluke 5522A with PQ Option	329.999 mV/0.9 A	$0.85 \times 10^{-3} P$	Laboratory/ Customer Premises
		329.999 mV/2.1999 A	$0.89 \times 10^{-3} P$	
		329.999 mV/2.2 A	$0.93 \times 10^{-3} P$	
		329.999 mV/4.4999 A	$1.3 \times 10^{-3} P$	
		329.999 mV/4.5 A	$1.3 \times 10^{-3} P$	
		329.999 mV/20.5 A	$1.7 \times 10^{-3} P$	
		330 mV/0.33 A	$1.1 \times 10^{-3} P$	
		330 mV/0.8999 A	$0.91 \times 10^{-3} P$	
		330 mV/0.9 A	$0.88 \times 10^{-3} P$	
		330 mV/2.1999 A	$0.92 \times 10^{-3} P$	
		330 mV/2.2 A	$0.97 \times 10^{-3} P$	
		330 mV/4.4999 A	$1.4 \times 10^{-3} P$	
		330 mV/4.5 A	$1.3 \times 10^{-3} P$	
		330 mV/20.5 A	$1.8 \times 10^{-3} P$	
		1020 V/0.33 A	$1.1 \times 10^{-3} P$	
1020 V/0.8999 A	$0.90 \times 10^{-3} P$			
1020 V/0.9A	$0.87 \times 10^{-3} P$			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
AC Power	Direct Method using Fluke 5522A with PQ Option	1020 V/2.1999 A	$0.91 \times 10^{-3} P$	Laboratory/ Customer Premises
		1020 V/2.2 A	$0.96 \times 10^{-3} P$	
		1020 V/4.4999 A	$1.4 \times 10^{-3} P$	
		1020 V/4.5 A	$1.3 \times 10^{-3} P$	
		1020 V/20.5 A	$1.8 \times 10^{-3} P$	
Oscilloscope	Direct Method using Fluke 5522A with SC1100 Option:			
	Relative Deviation Δy of the vertical Axis (measurement range):	2.5 mV to 6.6 V/ 50 Ω load, and 110 mV to 130 V/ 1 M Ω load at 1 kHz	$2.8 \times 10^{-3} U$	
	Oscilloscope Band Width	10 Hz to 1.1 GHz	$14 \times 10^{-3} f$	
Resistance Meters $\mu\Omega$; m Ω ; Ω ; k Ω ; M Ω	Direct Method using Decade Resistance Boxes: 50 $\mu\Omega$ to 2.0 Ω using Ductor Cal 5070 5.0 Ω to 3.0 M Ω using High Power Resistance Substituter HPRS-C-6-1	50 $\mu\Omega$;	$4.3 \times 10^{-3} R$	
		100 $\mu\Omega$;	$2.5 \times 10^{-3} R$	
		150 $\mu\Omega$;	$1.5 \times 10^{-3} R$	
		200 $\mu\Omega$	$1.4 \times 10^{-3} R$	
		0.5 m Ω	$12 \times 10^{-3} R$	
		1.0 m Ω	$5.9 \times 10^{-3} R$	
		1.5 m Ω	$3.9 \times 10^{-3} R$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Resistance Meters $\mu\Omega$; m Ω ; Ω ; k Ω ; M Ω	Direct Method using Decade Resistance Boxes: 50 $\mu\Omega$ to 2.0 Ω using Ductor Cal 5070 5.0 Ω to 3.0 M Ω using High Power Resistance Substituter HPRS-C-6-1	2.0 m Ω	$2.9 \times 10^{-3} R$	Laboratory/ Customer Premises
		5.0 m Ω	$1.2 \times 10^{-3} R$	
		10 m Ω	$0.72 \times 10^{-3} R$	
		15 m Ω	$0.58 \times 10^{-3} R$	
		20 m Ω	$0.58 \times 10^{-3} R$	
		50 m Ω	$0.16 \times 10^{-3} R$	
		100 m Ω	$0.13 \times 10^{-3} R$	
		150 m Ω	$0.12 \times 10^{-3} R$	
		200 m Ω	$0.13 \times 10^{-3} R$	
		0.5 Ω	$1.8 \times 10^{-3} R$	
		1.0 Ω	$0.89 \times 10^{-3} R$	
		1.5 Ω	$0.60 \times 10^{-3} R$	
		2.0 Ω	$0.46 \times 10^{-3} R$	
		5 Ω to 9 Ω	$1.9 \times 10^{-3} R$	
		10 Ω to 90 Ω	$1.9 \times 10^{-3} R$	
100 Ω to 900 Ω	$1.9 \times 10^{-3} R$			
1 k Ω to 9 k Ω	$1.9 \times 10^{-3} R$			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Resistance Meters $\mu\Omega$; m Ω ; Ω ; k Ω ; M Ω	Direct Method using Decade Resistance Boxes: 50 $\mu\Omega$ to 2.0 Ω using Ductor Cal 5070 5.0 Ω to 3.0 M Ω using High Power Resistance Substituter HPRS-C-6-1	100 Ω to 90 k Ω	$1.9 \times 10^{-3} R$	Laboratory/ Customer Premises
		100 k Ω to 900 k Ω	$1.9 \times 10^{-3} R$	
		1.0 M Ω	$17 \times 10^{-3} R$	
		2.0 M Ω	$17 \times 10^{-3} R$	
		3.0 M Ω	$12 \times 10^{-3} R$	
Insulation Resistance Tester	Direct Method using Decade Meg Ohm Box	0.1 M Ω to 9.99 M Ω	$2.3 \times 10^{-3} R$	
		10 M Ω to 99.9 M Ω	$8.5 \times 10^{-3} R$	
		100 M Ω to 1000 M Ω	$12 \times 10^{-3} R$	
Calibration of calibrators				
DC Voltage	Direct Method using Fluke 8846A <i>U = Measured Voltage value</i>	0 to 100 mV	$43 \times 10^{-6} U + 4 \mu V$	Laboratory/ Customer Premises
		>100mV to 1 V	$31 \times 10^{-6} U + 8 \mu V$	
		>1V to 10 V	$30 \times 10^{-6} U + 57 \mu V$	
		>10V to 100V	$46 \times 10^{-6} U + 0.69 mV$	
		>100V to 1000V	$49 \times 10^{-6} U + 12 mV$	
AC Voltage	Direct Method using Fluke 8846A <i>U = Measured Voltage value</i>	0 to 100 Mv		
		5 Hz	$4.1 \times 10^{-3} U + 46 \mu V$	
		>10 Hz to 20 kHz	$0.72 \times 10^{-3} U + 46 \mu V$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of calibrators				
AC Voltage	Direct Method using Fluke 8846A <i>U = Measured Voltage value</i>	0 to 100 Mv		Laboratory/ Customer Premises
		>20 kHz to 50 kHz	$1.5 \times 10^{-3} U + 58 \mu V$	
		>50 kHz to 100 kHz	$7 \times 10^{-3} U + 93 \mu V$	
		>100 mV to 1 V		
		5 Hz to 10 Hz	$4.1 \times 10^{-3} U + 0.35 mV$	
		>10 Hz to 20 kHz	$0.70 \times 10^{-3} U + 0.35 mV$	
		>20 kHz to 50 kHz	$1.4 \times 10^{-3} U + 0.58 mV$	
		>50 kHz to 100 kHz	$6.9 \times 10^{-3} U + 0.93 mV$	
		>1V to 10 V		
		5 Hz to 10 Hz	$4.1 \times 10^{-3} U + 3.5 mV$	
		10 Hz to 20 kHz	$0.7 \times 10^{-3} U + 3.5 mV$	
		20 kHz to 50 kHz	$1.4 \times 10^{-3} U + 5.8 mV$	
		50 kHz to 100 kHz	$7 \times 10^{-3} U + 9.3 mV$	
		>10 V to 100 V		
		5 Hz to 10 Hz	$4.1 \times 10^{-3} U + 35 mV$	
		10 Hz to 20 kHz	$0.7 \times 10^{-3} U + 35 mV$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of calibrators				
AC Voltage	Direct Method using Fluke 8846A <i>U = Measured Voltage value</i>	>10 V to 100 V		Laboratory/ Customer Premises
		20 kHz to 50 kHz	$1.4 \times 10^{-3} U + 58 \text{ mV}$	
		50 kHz to 100 kHz	$7 \times 10^{-3} U + 93 \text{ mV}$	
		>100 V to 1000 V		
		5 Hz to 10 Hz	$4.1 \times 10^{-3} U + 0.35 \text{ V}$	
		>10 Hz to 20 kHz	$0.72 \times 10^{-3} U + 0.35 \text{ V}$	
		>20 kHz to 50 kHz	$1.8 \times 10^{-3} U + 0.55 \text{ V}$	
		>50 kHz to 100 kHz	$7.0 \times 10^{-3} U + 0.92 \text{ V}$	
DC Current	Direct Method using Fluke 8846A <i>I = Measured Current value</i>	0 to 100 μA	$0.59 \times 10^{-3} I + 0.03 \mu\text{A}$	
		>100 μA to 1 mA	$0.58 \times 10^{-3} I + 0.06 \mu\text{A}$	
		>1 mA to 10 mA	$0.58 \times 10^{-3} I + 2.3 \mu\text{A}$	
		>10 mA to 100 mA	$0.58 \times 10^{-3} I + 5.8 \mu\text{A}$	
		>100 mA to 1 A	$0.59 \times 10^{-3} I + 0.23 \text{ mA}$	
		> 1 A to 10 A	$1.8 \times 10^{-3} I + 0.92 \text{ mA}$	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Electrical Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Calibration of calibrators				
AC Current	Direct Method using Fluke 8846A <i>I = Measured Current value</i>	0 to 100 μ A 10 Hz to 1 kHz	$1.2 \times 10^{-3} / + 0.05 \mu$ A	Laboratory/ Customer Premises
		>100 μ A to 1 mA 10 Hz to 1 kHz	$1.2 \times 10^{-3} / + 0.46 \mu$ A	
		>1 mA to 10 mA 10 Hz to 1 kHz	$1.2 \times 10^{-3} / + 4.6 \mu$ A	
		>10 mA to 100 mA 10 Hz to 1 kHz	$1.2 \times 10^{-3} / + 46 \mu$ A	
		>100 mA to 1 A 10 Hz to 1 kHz	$1.2 \times 10^{-3} / + 0.46$ mA	
		>1 A to 10 A 10 Hz to 1 kHz	$1.8 \times 10^{-3} / + 6.9$ mA	
Resistance	Direct Method using Fluke 8846A <i>R = Measured Resistance value</i>	0.0001 Ω to 10.0000 Ω	$0.12 \times 10^{-3} R + 3.5$ m Ω	
		10.001 Ω to 100.000 Ω	$0.12 \times 10^{-3} R + 4.6$ m Ω	
		0.1001 k Ω to 1.0 k Ω	$0.12 \times 10^{-3} R + 12$ m Ω	
		1.0001 k Ω to 10.0000 k Ω	$0.12 \times 10^{-3} R + 0.12$ Ω	
		10.001 k Ω to 100.000 k Ω	$0.12 \times 10^{-3} R + 1.2$ Ω	
		0.10001 M Ω to 1.00000 M Ω	$0.12 \times 10^{-3} R + 11.4$ Ω	
		1.0001 M Ω to 10.0000 M Ω	$0.47 \times 10^{-3} R + 0.12$ k Ω	
		10.001 M Ω to 100.000 M Ω	$9.3 \times 10^{-3} R + 12$ k Ω	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Electrical Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
---	--------------------	-------------------------	---	----------

Accreditation Scope

Force Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Force Verification /Calibration of Compression testing machines	GTS-WP-06 based on BS EN ISO 7500-1	50 kN up to 3000 kN	0.24% of reading using force transducer class 1, ISO 376	Client Premises
		150 kN up to 3000 kN	0,45 % of indicating reading using force transducer class 2, ISO 376	
Force Verification /Calibration of tensile testing machines	GTS-WP-06 based on BS EN ISO 7500-1	6,2 kN up to 300 kN	0.24 % of reading using force transducer class 1, ISO 376	Client Premises
Proving rings for soil testing apparatus	GTS-WP-08	400 N up to 50 kN	0.7 % of reading	Laboratory
Push-Pull gauge	GTS-WP-08	45 N up to 50 kN	0,3 % of reading	
Force gauge and load cell with indicator for industrial applications	GTS-WP-08	100 N up to 50 kN	1.0 % of reading	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Mass and Balance Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Conventional Mass (F1 , F2 Class) Up to 5 kg. M Class for 10 & 20 kg	GTS-WP-17 Substitution Weighing with air buoyancy Error ABBA weighing cycle based on OIML R-111:2004, OIML- D28:2004, PTB-Guide MA- 40	1 mg	0.02 mg	Laboratory
		2 mg	0.02 mg	
		5 mg	0.02 mg	
		10 mg	0.02 mg	
		20 mg	0.02 mg	
		50 mg	0.02 mg	
		100 mg	0.02 mg	
		200 mg	0.02 mg	
		500 mg	0.02 mg	
		1 g	0.02 mg	
		2 g	0.02 mg	
		5 g	0.03 mg	
		10 g	0.03 mg	
		20 g	0.05 mg	
		50 g	0.09 mg	
		100 g	0.19 mg	
200 g	0.35 mg			
500 g	0.81 mg			

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Mass and Balance Calibration
004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Conventional Mass (F1 , F2 Class) Up to 5 kg. M Class for 10 & 20 kg	GTS-WP-17 Substitution Weighing with air buoyancy Error ABBA weighing cycle based on OIML R-111:2004, OIML- D28:2004, PTB-Guide MA- 40	1 kg	1.6 mg	Laboratory
		2 kg	3.4 mg	
		5 kg	8.4 mg	
		10 kg	0.16 g	
		20 kg	0.17 g	
Calibration of top loading direct reading weighing balance	GTS-WP-01 Based on the requirements of ASTM E 898 Calibrated weights - E1 ,E2,F1, F2 & M1 weights	Up to 100 g	0.0001 g	Laboratory/ Customer Premises
		Up to 210 g	0.0002 g	
		Up to 0.5 kg	0.0006 g	
		Up to 1 kg	0.001 g	
		Up to 5 kg	0.009 g	
		Up to 10 kg	0.013 g	
		Up to 30 kg	0.23 g	
		Up to 100 kg	0.45 g	
		Up to 500 kg	2.27 g	
		Up to 1000 kg	0.46 kg	
Calibration concrete and asphalt batching plants (Hopper Scale)	Hopper Scale calibration of concrete and asphalt batching plants ASTM C94/C94M & NIST Handbook 44	Up to 5000 kg	0.05%	Customer Premises

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Pressure Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Pneumatic Pressure Calibration of Digital & Analogue Pressure Gauges	GTS-WP-02 Based on the requirements of BS EN 837-1 : 1998 Using Druck DPI 610, DPI 620	up to 20 bar	0.02%	Laboratory/ Customer Premises
Pneumatic Pressure Transducers, Transmitters and Switches	GTS-WP-02 Using Druck DPI 610, DPI 104 and Fluke 8846 multimeter	Up to 100 bar	0.11%	Laboratory/ Customer Premises
Vacuum gauge calibration –Analogue & Digital	GTS-WP-03 Based on the requirements of BS EN 837-1 : 1998 and ISO/TS 3567 Using Druck DPI 610, DPI 620	0 up to – 1 bar	1 mbar	Laboratory/ Customer Premises
Hydraulic Pressure Transducers, Transmitters and Switches	GTS-WP-02 Using Budenberg DWT 580HX Piston Cylinder 030L & Fluke 8846 multimeter	Up to 1200 bar	0.11%	Lab
Hydraulic pressure Digital & Analogue Pressure Gauges & pressure modules	GTS-WP-02 Using Budenberg DWT 580HX Piston Cylinder 030L	Up to 1200 bar	0.02%	Lab

Calibration of hydraulic pressure balance	GTS-WP-143 based on OIML R110 and EURAMET cg-3 Version 1.0 (03/2011)	up to 1400 bar	0.01%	Lab
Calibration of pneumatic pressure balance	GTS-WP-143 based on OIML R110 and EURAMET cg-3 Version 1.0 (03/2011)	0.5 to 25 bar	0.01%	Lab
Calibration of Mercury and dial Sphygmomanometer	GTS-WP-184	0 mm Hg to 350 mm Hg	0.58 % rdg.	Laboratory/ Customer Premises

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Temperature and Humidity Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Liquid-in-glass thermometers	GTS-WP-13 Based on BS 1041-2-1	-30 °C up to 150 °C	0.16 °C	Laboratory
Direct reading thermometers	GTS-WP-15 (in house method)	-40 °C up to 160 °C	0.16 °C	
		>160 °C up to 500 °C	0.4 °C	
		>500 °C up to 900 °C	1.3 °C	
		>900 °C up to 1200 °C	4 °C	
Dial Thermometers	GTS-WP-14 Based on EN 13190	-30°C up to 160°C	0.16 °C	
		>160°C up to 400°C	2.0 °C	
		>400°C up to 800°C	4.0 °C	
Base Metal Thermocouples	GTS-WP-12	-40 °C up to 250 °C	0.3 °C	
		>250 °C up to 600 °C	0.6 °C	
		>600 °C up to 900 °C	0.9 °C	
		>900 °C up to 1200 °C	4.0 °C	
Noble Metal Thermocouple	GTS-WP-12	0°C up to 600°C	0.6°C	
		>600°C up to 900°C	0.8°C	
		>900°C up to 1200°C	1.7°C	
Climatic Chamber	GTS-09 Based on DKD-R-5- 7 (9 points)	30°C up to 180°C	1.1°C	Laboratory/ Client Premises

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Temperature and Humidity Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Water Bath, incubators	GTS-09 Based on DKD-R-5-7 (5 points)	5°C up to 95°C	1.1°C	Laboratory/ Client Premises
Freezer/Chiller	GTS-154 Based on DKD-R-5-7 (9 points)	-30°C up to 95°C	1.1°C	
Furnace, Oven	GTS-WP-09 Based on DKD-R5-7(9 points, muffle furnace: 1 point)	30°C up to 180°C	1.1°C	
		>180°C up to 300°C	1.3°C	
		>300°C up to 800°C	4.0°C	
		>800°C up to 1200°C	9.0°C	
Auto Clave(Temperature)	GTS-WP-155 Based on DKD-R5-7 (5 – 9 points)	100°C up to 140°C	0.8 °C	
Refrigerator	GTS-WP-176	- 40°C to 20°C	0.8 °C	
Stirred Liquid bath	GTS-WP-182	-35°C to 165°C	0.6 °C	
		>165°C to 300°C	0.7 °C	
Dry Block Calibrator	GTS-WP-177	Atmospheric temp. to 250°C	0.4 °C	
		>250 to 400°C	0.6 °C	
		>400 to 650°C	0.8 °C	
		>650°C to 900°C	1.2 °C	
		900°C to 1100°C	2.5 °C	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope
Temperature and Humidity Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Humidity meter / Transmitter	GTS-WP-178	10% of RH to 90 % of RH	1.0% of RH	Laboratory
RTD with/ without Temperature Indicator	GTS-WP-179	-45°C to 40°C	0.22 °C	Laboratory
		>40°C to 200°C	0.37 °C	
		>200°C to 600°C	0.45 °C	
Infrared Thermometer	GTS-WP - 150 Comparison method	-30°C to 0°C	3.5 °C	Laboratory/ Client Premises
		>0°C to 600°C	4.0°C	
Temperature Transducer/ Transmitter/ Switch	GTS-WP-181	-30 °C to 150°C	0.8 °C	Laboratory/ Client Premises
		>150°C to 850°C	0.9°C	
Data Logger (Temperature, Humidity)	GTS-WP-183	-10 °C to 70°C	0.8 °C	Laboratory
		10% to 90 % of RH	0.9% of RH	

Accreditation Scope

Torque Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 02

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Torque Hand Torque Tools	GTS-WP-31 based on: ISO 6789-1: 2017 and ISO 6789-2: 2017	0,5 N·m to 340 N·m Torque transducers TTC 400	1,0 % of reading	Laboratory
		81 N·m to 813 N·m Torque transducer TTC 12	1,0 % of reading	
		271 N·m to 2711 N·m Torque transducer TTC 14	1,0 % of reading	
Torque Transducers	GTS-WP-31 based on BS 7882:2017	0.45 N·m to 5.65 N·m	0.5 % of reading	
		3.39 N·m to 45.19 N·m	0.4 % of reading	
		9.03 N·m to 112.98 N·m	0.3 % of reading	
		27.12 N·m to 338.96 N·m	0.3 % of reading	
		81.35 N·m to 813.49 N·m	0.3 % of reading	
		271 N·m to 2711 N·m	0.8 % of reading	

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

Accreditation Scope

Volume Calibration

004-LB-CAL

General Const. Lab Calibration LLC

Industrial Area # 3, Sharjah-United Arab Emirates

Issue no.: 08

Date: 25-12-2019

Valid to: 25-05-2021

Calibration Field/ Measuring Quality	Calibration Method	Range and Specification	Calibration Measurement Capability (CMC)*	Location
Pipette Fixed and Variable Volume Pipette	GTS-WP-61A & GTS-WP-61 Gravimetric method ISO 8655-6 & ISO 4787	0.5 µl to 100 µl	0.57 µl	Laboratory
		>100 µl to 2000 µl	1.30 µl	
		>2000 µl to 5000 µl	1.5 µl	
		> 5 ml to 100 ml	0.11%	
Beaker	GTS-WP-61 & 61B Gravimetric method ISO 8655-6 & ISO 4787	50 ml to 5000 ml	0.70%	Laboratory
Graduated cylinder		50 ml to 5000 ml	0.70%	
Volumetric Measuring Flask		> 5 ml to 100 ml	0.10%	
Specific Gravity Bottle		> 100 ml to 5000 ml	0.04%	
		5 ml to 100 ml	0.03%	
Volume Jar & Prover Tanks	GTS-WP-61B Gravimetric method ISO 8655-6 & ISO 4787	>5 L to 20 L	0.02%	Laboratory

* Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.